



Figure S2. Testing of all possible primer combination of the MPXV-RPA-assay. All nine primer combination were tested with the molecular DNA standard with a concentration of 10^5 DNA molecules/ μ l. A mixing step was conducted after 230 sec. The combination FP3 + RP3 showed the earliest and highest fluorescence signal and was therefore chosen for further assay validation.